



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Load Cell
Tension & Compression
Model: TS615, TS60063, TS60001 & TS60050
 n_{max} : 4500, Class III, Single Cell, 100 to 500 lb
5500, Class III, Single Cell, 1000 to 20 000 lb
Capacity: 100 to 20 000 lb
Accuracy Class: III

Submitted By:

Transcell Technology, Inc.
975 Deerfield Parkway
Buffalo Grove, IL 60089
Tel: 847-419-9180
Fax: 847-419-1515
Contact: Jon Heinlein
Email: jheinlein@transcell.com
Web site: www.transcell.com

Standard Features and Option

The specific load cell capacities, v_{min} values, and minimum dead loads covered by this Certificate are listed in the table below.
Nominal output: 2.0 and 3.0 mV/V
Steel Stainless and Alloy Steel material
4 wire design
Minimum Dead Load: 0 lb

Models	Capacity	v_{min} Class III Single cell	n_{max} Class III Single cell
TS615, TS60063, TS60050 & TS60001 Load Cells Tested: 200 kg & 2000 kg	100 lb	0.0067 lb	4500
	250 lb	0.017 lb	4500
	500 lb	0.03 lb	4500
	1000 lb	0.07 lb	5500
	2000 lb	0.14 lb	5500
	3000 lb	0.21 lb	5500
	5000 lb	0.36 lb	5500
	10 000 lb	0.70 lb	5500
	15 000 lb	1.07 lb	5500
	20 000 lb	1.43 lb	5500

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Kristin Maëy
Chairman, NCWM, Inc.

Jerry Buendel
Chairman, National Type Evaluation Program Committee

Issued: August 24, 2016

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.

**Transcell Technology Inc.****Load Cell / TS615, TS60063, TS60001 & TS60050**

Application: The load cells may be used in Class III scales for single cell and multiple cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the v_{\min} value, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions (n_{\max}) and with greater v_{\min} values than those listed on the certificate. However, the load cells must be marked with the appropriate n_{\max} and v_{\min} for which the load cell may be used.

Identification: A pressure sensitive identification label located on the cell, states manufacturer name, model, serial number and rated capacity. Other pertinent information will be specified on the Calibration Certificate accompanying the cell.

Test Conditions: This Certificate is issued based upon the following tests and upon information provided by the manufacturer. A 200 kg and 2000 kg capacity load cells were tested by the NMi Certain B.V. at The Netherlands facility. Testing was conducted in accordance with the OIML DoMC Mutual Acceptance Arrangement, signed by the NCWM as a utilizing participant for load cell testing. Testing was conducted using deadweights as the reference standard. The load cells were tested over a temperature range of -10 °C to 40 °C with tests run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test to determine sensitivity of the load cell design to changes in barometric pressure was conducted. The data were analyzed for single load cell applications. OIML R60 selection criteria were used to determine cells tested.

Evaluated By: C. Bontenbal (NMI), A. Tjoa (NMI), M.M.J. Meijer (NMI), E. van der Grinten (NMI)

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2016. NCWM, Publication 14: Weighing Devices, 2016.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM)

Example of Device: